

Mouse uPA Activity Assay

For Research Use Only

INTRODUCTION

Urokinase-type Plasminogen Activator (uPA) is a glycosylated serine protease that is synthesized in endothelial and kidney epithelial cells. There are two forms of uPA, the low molecular weight (LMW ~31 kDa) and high molecular weight (HMW ~55 kDa). The HMW form undergoes autoproteolysis resulting in the LMW plus an 18.5 kDa amino terminal fragment (ATF). It is this ATF that has been shown to inhibit proliferation and invasion of cancer cells by binding to uPA receptors(1).

PRINCIPLES OF PROCEDURE

This is an ELISA (Enzyme-Linked Immunosorbent Assay) for the quantitative analysis of active uPA levels in biological fluid. This test kit operates on the basis of sandwich ELISA where active uPA enzyme complexes with PAI-1 and is quantitated with the use of an HRP labeled secondary antibody.

First the biotinylated PAI-1 binds to the avidin coated wells. Next, active uPA, present in the standard or unknown samples, complexes with PAI-1. Inactive or complexed uPA is removed in a subsequent wash step. A primary antibody specific for uPA is then added to each well followed by the HRP conjugated secondary antibody. The bound conjugated secondary antibody is detected by the addition of substrate, which generates an optimal color after 10 minutes. Quantitative test results may be obtained by the measure and comparison of the sample and standard absorbance readings when read with a microplate reader at 450 nm.

MATERIALS PROVIDED

Component	Contents	Quantity	Storage	Cat. No.
Coated Plate	Avidin coated 96-well plate	1 plate	4°C	PL92a
Standard	Mouse uPA activity standard (lyophilized)	1 vial	4°C	PL92b
Biotinylated PAI-1	Biotinylated PAI-1 (lyophilized)	1 vial	4°C	PL92c
Primary Antibody	Anti-mouse uPA antibody (lyophilized)	1 vial	4°C	PL92d
10x Wash Buffer	10x solution for washing plate	50 mL	4°C	PL92e
TMB Substrate	TMB Substrate	10 mL	4°C	PL92f
Secondary Antibody	Anti-rabbit HRP conjugated antibody	1 vial	4°C	PL92g
10x TBS Solution	Tris Buffered Saline; pH 7.4	5 mL	4°C	PL92h

MATERIALS NEEDED BUT NOT PROVIDED

- 1 N H₂SO₄
- TBS Buffer (see Reagent Preparation)
- Blocking Buffer (see Reagent Preparation)
- DI Water
- Microplate reader with 450 nm filter
- Microplate shaker with uniform horizontal circular movement up to 300 rpm
- Precision pipettes that range from 10 µL-1000 µL and disposable tips

STORAGE

1. Store the kit and all of its components at 4°C before use.

2. If not using the entire plate at once, prepare only the appropriate amount of Primary Antibody, Biotinylated PAI-1, and uPA Standard. The remaining stock solutions should be aliquoted, frozen and stored at -70°C . **DO NOT** freeze/thaw the Primary Antibody or Standard more than once. All other components should remain refrigerated. Store unused portions of the microplate in a pouch with a desiccant at 4°C .

WARNINGS AND PRECAUTIONS

1. This kit is designed to work properly as provided and instructed. Additions, deletions or substitutions to the procedure or reagents are not recommended, as they may be detrimental to the assay.
2. Exercise universal precautions during the performance or handling of this kit or any component contained therein.

PROCEDURAL NOTES

1. This assay should be run at room temperature.
2. Use aseptic technique when opening and dispensing reagents.
3. To minimize error due to handling, wipe the exterior bottom of the microplate wells with a lint-free paper towel.

SAMPLE COLLECTION, STORAGE, AND PREPARATION

Samples should be collected using 0.01 M trisodium citrate, acidified citrate or Stabilyte™ (Biopool, cat# 102080) collection media. Collection should be in accordance with the collection vial manufacturer's instructions or in a 1:10 ratio of collection media to blood.

Immediately upon collection of blood, the samples should be centrifuged at $3000 \times g$ for 15 minutes. This should ensure the removal of platelets as they can release PAI-1 that, in turn, complexes with uPA. The plasma can be transferred to a clean plastic tube and stored frozen for up to one month. Samples are stable for approximately 24 hours when stored at 4°C with the Sabilyte™ media or one month if stored at -20°C .

Note: Detergents such as Triton X cause interference with the assay. If using detergent extracted samples, it is necessary to dialyze the samples overnight to remove the detergent.

REAGENT PREPARATION

The following solutions should be prepared fresh before starting the assay.

1. **TBS Buffer:** 0.1 M TRIS, 0.15 M NaCl, pH 7.4
2. **3% BSA Blocking Buffer:** 3% BSA in TBS Buffer.
3. **10x Wash Buffer:** Dilute the 50 mL of concentrate to 1x with 450 mL of DI water prior to use.
4. **Biotinylated PAI-1:** Reconstitute with 3% BSA Blocking Buffer as directed on the vial and vortex gently to mix. Prepare immediately prior to use.
5. **Standard:** Reconstitute with 3% BSA Blocking Buffer as directed on the vial and vortex gently to mix. Prepare according to included Standard Dilution Table immediately prior to use.
6. **Primary Antibody:** Reconstitute with 3% BSA Blocking Buffer as directed on the vial and vortex gently to mix. Prepare immediately prior to use.
7. **Secondary Antibody:** Dilute with 3% BSA Blocking Buffer as directed on the vial and vortex gently to mix. Prepare immediately prior to use.

ASSAY PROCEDURE

1. Add $100 \mu\text{l}$ of reconstituted Biotinylated PAI-1 to all of the wells. Shake the plate at 300 rpm for 30 minutes at room temperature (RT).
2. Wash the plate 3 times according to the following wash procedure:
 - a. Remove the contents of each well by inversion of the plate.
 - b. Tap out the remaining contents of the plate onto a lint free paper towel.
 - c. Add $300 \mu\text{L}$ of 1x Wash Buffer.
 - d. Let stand for 2-3 minutes.

- e. Remove the contents of each well by inversion of plate into an appropriate disposal device.
 - f. Repeat procedure two more times, then proceed to step “g”.
 - g. Tap out the remaining contents of the plate onto a lint free paper towel, then proceed to step 3.
3. Add 100 μ l of the Standards and unknowns to the wells in duplicate. Dilutions of the unknown may be made in 3% BSA Blocking Buffer. For a suggested plate layout, see Scheme I below. Shake the plate at 300 rpm for 30 minutes at RT.

NOTE: If using citrated plasma or samples with a pH lower than 6.0, add 30 μ L of the 10x TBS Solution to the well prior to adding the sample. This step is unnecessary if the pH is neutral.
 4. Wash the plate three times as in step 2.
 5. Add 100 μ l of the Primary Antibody to each well. Shake the plate at 300 rpm for 30 minutes at RT.
 6. Wash the plate three times as in step 2.
 7. Add 100 μ l of the Secondary Antibody to each well. Shake the plate at 300rpm for 30 minutes at RT.
 8. Wash the plate three times as in step 2.
 9. Add 100 μ l of TMB Substrate to each well. Shake the plate at 300 rpm for 2-10 minutes at RT.
 10. Stop the reaction by adding 50 μ l of 1N H₂SO₄ to each well and read the plate at 450 nm.


Scheme I:

	1	2	3	4	5	6	7	8	9	10	11	12
A	S ₀	S ₁	S ₂	S ₃	S ₄	S ₅	S ₆	S ₇	S ₈	S ₉	U ₁	U ₂
B	S ₀	S ₁	S ₂	S ₃	S ₄	S ₅	S ₆	S ₇	S ₈	S ₉	U ₁	U ₂
C	U ₃	U ₄	U ₅	U ₆	U ₇	U ₈	U ₉	U ₁₀	U ₁₁	U ₁₂	U ₁₃	U ₁₄
D	U ₃	U ₄	U ₅	U ₆	U ₇	U ₈	U ₉	U ₁₀	U ₁₁	U ₁₂	U ₁₃	U ₁₄
E	U ₁₅	U ₁₆	U ₁₇	U ₁₈	U ₁₉	U ₂₀	U ₂₁	U ₂₂	U ₂₃	U ₂₄	U ₂₅	U ₂₆
F	U ₁₅	U ₁₆	U ₁₇	U ₁₈	U ₁₉	U ₂₀	U ₂₁	U ₂₂	U ₂₃	U ₂₄	U ₂₅	U ₂₆
G	U ₂₇	U ₂₈	U ₂₉	U ₃₀	U ₃₁	U ₃₂	U ₃₃	U ₃₄	U ₃₅	U ₃₆	U ₃₇	BLK
H	U ₂₇	U ₂₈	U ₂₉	U ₃₀	U ₃₁	U ₃₂	U ₃₃	U ₃₄	U ₃₅	U ₃₆	U ₃₇	BLK

CALCULATIONS

1. Average the O.D. values for each pair of duplicate wells.
2. Plot the A₄₅₀ against the concentration of uPA in the standards.
3. Fit a straight line through the points using a linear fit procedure.
4. Calculate the uPA concentrations in the unknowns using the standard curve.

Typical Standard Curve:



PERFORMANCE CHARACTERISTICS

Assay Range: 0.125-10 ng/mL

Samples with uPA levels higher than 10 ng/mL should be diluted in similar media devoid of active uPA or 3% BSA Blocking Buffer.

REFERENCES

1. Declerck, P., *et al.*, (1995) *Thromb Haemostas.*, Novu 74(5):1305-1309
2. Singh, I., *et al.*, (2003) *Circulation* 107(6):869-875
3. Kjølner, L., (2002) *Biol. Chem.* 383:5-19
4. Yang, Y., *et al.*, (2001) *J. Immunol.* 167(2):1047-1052

DISCLAIMER

This information is believed to be correct but does not purport to be all-inclusive and shall be used only as a guide. Oxford Biomedical Research, Inc. shall not be held liable for any damage resulting from handling or from contact with the above product. See catalog for additional terms and conditions of sale.

ORDERING INFORMATION

For additional kits or a complete catalog please call 800-692-4633.

TECHNICAL SUPPORT

If you need technical information or assistance with assay procedures, call our Technical Support Department at 800-692-4633 or 248-852-8815. Our staff will be happy to answer your questions about this or any other product in the Oxford Biomedical line.

GUARANTEE AND LIMITATION OF REMEDY

Oxford Biomedical Research, Inc. makes no guarantee of any kind, expressed or implied, which extends beyond the description of the material in this ELISA kit, except that these materials and this kit will meet our specifications at the time of delivery. Buyer's remedy and Oxford Biomedical Research, Inc.'s sole liability hereunder is limited to, at Oxford Biomedical Research, Inc.'s option, refund of the purchase price of, or the replacement of, material that does not meet our specification. By acceptance of our products, Buyer indemnifies and holds Oxford Biomedical Research, Inc. harmless against, assumes all liability for the consequence of its use or misuse by the Buyer, its employees, or others. Said refund or replacement is conditioned of Buyer notifying Oxford Biomedical Research, Inc. within (30) days of the receipt of product. Failure of Buyer to give said notice within said thirty (30) days shall constitute a waiver by the Buyer of all claims hereunder with respect to said material(s).

Oxford Biomedical Research, Inc.
P.O. Box 522
Oxford, MI 48371 U.S.A.

Orders: 800-692-4633
Technical Service: 248-852-8815
Fax: 248-852-4466
E-mail: info@oxfordbiomed.com

Made in the U.S.A.